

LOT	B255L
	2025-08-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry


The barcodes may be read in any order.

If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.6	2.2	2.2	2.10
RBC	10 ⁶ /μL	±0.15	2.34	2.34	2.31
HGB	g/dL	±0.4	6.0	6.2	6.16
HCT	%	±2.0	18.6	18.6	20.7
MCV	fL	±9.0	79.3	79.3	89.4
MCH	pg	±2.4	25.6	26.5	26.7
MCHC	g/dL	±3.0	32.3	33.3	29.8
PLT	10 ³ /μL	±30	67	67	79.6
MPV	fL	±3.0	5.6	8.1	8.2
RDW-CV *	%	±3.0	18.2	18.2	16.6
LY%	%	±10.0	54.9	54.9	56.35
MO%	%	±5.0	7.6	7.6	4.06
GR%	%	±10.0	37.5	37.5	39.59
LY	10 ³ /μL	±0.5	1.2	1.2	1.18
MO	10 ³ /μL	±0.2	0.2	0.2	0.09
GR	10 ³ /μL	±0.5	0.8	0.8	0.83

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.

MEK-1305
QC ESR
45±17 mm

Manual Registration	
Lot	255
Level	Low
Control	MEK-3D



<https://www.nihonkohden.com/spi/index.html>



NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



MEK-3DL



B255L



2025-08-05



0694-901684T

