


LOT	B258H
	2025-11-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±2.2	21.3	21.3	21.06
RBC	10 ⁶ /μL	±0.25	5.90	5.90	5.86
HGB	g/dL	±0.7	18.8	18.8	19.09
HCT	%	±4.0	54.5	54.5	61.9
MCV	fL	±7.0	92.3	92.3	105.7
MCH	pg	±2.8	31.9	31.9	32.6
MCHC	g/dL	±3.0	34.5	34.5	30.8
PLT	10 ³ /μL	±70	519	519	540.8
MPV	fL	±3.0	5.6	8.2	8.5
RDW-CV *	%	±3.0	16.5	16.5	17.6
LY%	%	±10.0	17.1	17.1	14.48
MO%	%	±5.0	1.2	1.2	1.33
GR%	%	±10.0	81.7	81.7	84.19
LY	10 ³ /μL	±1.8	3.6	3.6	3.05
MO	10 ³ /μL	±0.7	0.3	0.3	0.28
GR	10 ³ /μL	±2.0	17.4	17.4	17.73

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.

Manual Registration	
Lot	258
Level	High
Control	MEK-3D



<https://www.nihonkohden.com/spi/index.html>



NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN

