

LOT	B25YH
	2026-02-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±2.2	21.3	21.3	20.89
RBC	10 ⁶ /μL	±0.25	5.98	5.98	5.95
HGB	g/dL	±0.7	19.2	19.2	19.37
HCT	%	±4.0	55.6	55.6	63.4
MCV	fL	±7.0	92.9	92.9	106.5
MCH	pg	±2.8	32.1	32.1	32.6
MCHC	g/dL	±3.0	34.5	34.5	30.6
PLT	10 ³ /μL	±70	515	515	537.6
MPV	fL	±3.0	5.4	8.1	8.4
RDW-CV *	%	±3.0	16.5	16.5	17.6
LY%	%	±10.0	14.5	14.5	12.43
MO%	%	±5.0	1.2	1.2	1.43
GR%	%	±10.0	84.3	84.3	86.14
LY	10 ³ /μL	±1.8	3.1	3.1	2.60
MO	10 ³ /μL	±0.7	0.3	0.3	0.30
GR	10 ³ /μL	±2.0	17.9	17.9	17.99

Manual Registration	
Lot	25Y
Level	High
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



MEK-3DH



B25YH



2026-02-05



0694-901686V