MULTI SERVICIOS LAB

INSERTOS

NIHON KOHDEN

MEK-3DL HEMATOLOGY CONTROL 2.0mL

ASSAY SHEET

LOT	B25YL
⌛	2026-02-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.

If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1

2

3

		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	$10^3/\mu L$	±0.6	2.1	2.1	2.07
RBC	$10^6/\mu L$	±0.15	2.43	2.43	2.36
HGB	g/dL	±0.4	6.0	6.2	6.16
HCT	%	±2.0	18.7	18.7	20.8
MCV	fL	±9.0	77.0	77.0	88.0
MCH	pg	±2.4	24.7	25.5	26.1
MCHC	g/dL	±3.0	32.1	33.2	29.6
PLT	$10^3/\mu L$	±30	57	57	67.3
MPV	fL	±3.0	5.4	8.1	7.6
RDW-CV *	%	±3.0	17.5	17.5	15.7
LY%	%	±10.0	57.7	57.7	56.25
MO%	%	±5.0	10.0	10.0	6.67
GR%	%	±10.0	32.3	32.3	37.08
LY	$10^3/\mu L$	±0.5	1.2	1.2	1.16
MO	$10^3/\mu L$	±0.2	0.2	0.2	0.14
GR	$10^3/\mu L$	±0.5	0.7	0.7	0.77

MEK-1305
QC ESR
45±17 mm

Manual Registration	
Lot	25Y
Level	Low
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.

https://www.nihonkohden.com/spi/index.html

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN

MEK-3DL

B25YL

2026-02-05

CE 0123 IVD

0694-901684V