

LOT	B262L
	2026-05-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.6	2.2	2.2	2.27
RBC	10 ⁶ /μL	±0.15	2.37	2.37	2.31
HGB	g/dL	±0.4	5.8	6.0	6.05
HCT	%	±2.0	18.2	18.2	20.1
MCV	fL	±9.0	76.9	76.9	87.1
MCH	pg	±2.4	24.5	25.3	26.2
MCHC	g/dL	±3.0	31.9	33.0	30.1
PLT	10 ³ /μL	±30	65	65	74.5
MPV	fL	±3.0	5.8	8.3	8.1
RDW-CV *	%	±3.0	18.1	18.1	16.4
LY%	%	±10.0	55.1	55.1	59.99
MO%	%	±5.0	6.4	6.4	3.08
GR%	%	±10.0	38.5	38.5	36.93
LY	10 ³ /μL	±0.5	1.2	1.2	1.36
MO	10 ³ /μL	±0.2	0.1	0.1	0.07
GR	10 ³ /μL	±0.5	0.9	0.9	0.84

MEK-1305
QC ESR
49±17 mm

Manual Registration	
Lot	262
Level	Low
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU

TOKYO 161-8560 JAPAN



MEK-3DL



B262L



2026-05-05



0694-901684W