

<b>LOT</b>	B262N
	2026-05-05
<b>OPEN</b>	14DAYS

**MEK-1301,1302,1305 Barcode Entry**

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1

2

3

		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
<b>WBC</b>	10 <sup>3</sup> /μL	±0.8	7.6	7.6	7.61
<b>RBC</b>	10 <sup>6</sup> /μL	±0.20	4.80	4.80	4.75
<b>HGB</b>	g/dL	±0.5	13.6	13.7	13.85
<b>HCT</b>	%	±3.0	41.0	41.0	46.3
<b>MCV</b>	fL	±7.0	85.5	85.5	97.5
<b>MCH</b>	pg	±2.4	28.3	28.5	29.2
<b>MCHC</b>	g/dL	±3.0	33.2	33.4	29.9
<b>PLT</b>	10 <sup>3</sup> /μL	±50	236	236	249.9
<b>MPV</b>	fL	±2.0	5.7	8.2	8.2
<b>RDW-CV *</b>	%	±3.0	16.7	16.7	17.0
<b>LY%</b>	%	±10.0	27.7	27.7	31.24
<b>MO%</b>	%	±5.0	2.6	2.6	2.50
<b>GR%</b>	%	±10.0	69.7	69.7	66.26
<b>LY</b>	10 <sup>3</sup> /μL	±1.0	2.1	2.1	2.38
<b>MO</b>	10 <sup>3</sup> /μL	±0.4	0.2	0.2	0.19
<b>GR</b>	10 <sup>3</sup> /μL	±1.0	5.3	5.3	5.04

<b>MEK-1305</b>
<b>QC ESR</b>
<b>10±5 mm</b>

Manual Registration	
Lot	<b>262</b>
Level	<b>Normal</b>
Control	<b>MEK-3D</b>

\* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>

**NIHON KOHDEN CORPORATION**  
1-31-4 NISHIOCHIAI, SHINJUKU-KU  
TOKYO 161-8560 JAPAN

MEK-3DN

B262N

2026-05-05

0694-901685W