

LOT	B265H
	2026-08-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±2.2	20.2	20.2	19.89
RBC	10 ⁶ /μL	±0.25	5.89	5.89	5.90
HGB	g/dL	±0.7	19.1	19.0	19.17
HCT	%	±4.0	55.0	55.0	62.8
MCV	fL	±9.0	93.3	93.3	106.5
MCH	pg	±2.8	32.4	32.3	32.5
MCHC	g/dL	±3.0	34.7	34.5	30.5
PLT	10 ³ /μL	±70	499	499	497.1
MPV	fL	±3.0	5.3	7.9	8.1
RDW-CV *	%	±3.0	16.3	16.3	17.8
LY%	%	±10.0	15.8	15.8	13.30
MO%	%	±5.0	1.0	1.0	1.15
GR%	%	±10.0	83.2	83.2	85.55
LY	10 ³ /μL	±1.8	3.2	3.2	2.65
MO	10 ³ /μL	±0.7	0.2	0.2	0.23
GR	10 ³ /μL	±2.0	16.8	16.8	17.01

Manual Registration	
Lot	265
Level	High
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



MEK-3DH



B265H



2026-08-05



0694-901686X