


LOT	B265L
	2026-08-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.



		RANGE	MEK -6400	MEK -6500	MEK -1301
			-6410	-6510	-1302
			-6420	-6550	-1305
			-6450		
WBC	10 ³ /μL	±0.6	1.8	1.8	1.81
RBC	10 ⁶ /μL	±0.15	2.42	2.42	2.36
HGB	g/dL	±0.4	5.9	6.0	6.04
HCT	%	±2.0	18.2	18.2	19.7
MCV	fL	±9.0	75.2	75.2	83.6
MCH	pg	±2.4	24.4	24.8	25.6
MCHC	g/dL	±3.0	32.4	33.0	30.7
PLT	10 ³ /μL	±30	58	58	65.9
MPV	fL	±3.0	6.0	8.3	8.2
RDW-CV *	%	±3.0	18.1	18.1	16.8
LY%	%	±10.0	54.3	54.3	59.13
MO%	%	±5.0	6.5	6.5	3.21
GR%	%	±10.0	39.2	39.2	37.66
LY	10 ³ /μL	±0.5	1.0	1.0	1.07
MO	10 ³ /μL	±0.2	0.1	0.1	0.06
GR	10 ³ /μL	±0.5	0.7	0.7	0.68

MEK-1305
QC ESR
48±17 mm

Manual Registration	
Lot	265
Level	Low
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN

