

LOT	B265N
	2026-08-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.8	7.6	7.6	7.61
RBC	10 ⁶ /μL	±0.20	4.77	4.77	4.78
HGB	g/dL	±0.5	14.0	14.1	14.16
HCT	%	±3.0	41.4	41.4	47.1
MCV	fL	±9.0	86.7	86.7	98.6
MCH	pg	±2.4	29.4	29.6	29.6
MCHC	g/dL	±3.0	33.8	34.1	30.1
PLT	10 ³ /μL	±50	240	240	247.5
MPV	fL	±2.0	5.4	8.0	7.9
RDW-CV *	%	±3.0	16.8	16.8	17.1
LY%	%	±10.0	30.2	30.2	32.12
MO%	%	±5.0	3.1	3.1	3.21
GR%	%	±10.0	66.7	66.7	64.67
LY	10 ³ /μL	±1.0	2.3	2.3	2.44
MO	10 ³ /μL	±0.4	0.2	0.2	0.24
GR	10 ³ /μL	±1.0	5.1	5.1	4.93

MEK-1305
QC ESR
10±5 mm

Manual Registration	
Lot	265
Level	Normal
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



MEK-3DN



B265N



2026-08-05



0694-901685X